



«TASDIQLASHMAN»

Bosh shifokor

[Handwritten signature]

Rentgen nurlari shaffof mobil stol

RENTGEN NURLARI SHAFFOF MOBIL STOL
Maqsad Bemorlarni yotqizish va rentgen diagnostikasini o'tkazish uchun
STOL O'LCHAMLARI
Stol usti rentgen shaffof materialdan yasalgan bo'lishi kerak
Stol tagining uzunligi 200 sm
Stol tagining kengligi 65 sm
Maksimal ruxsat etilgan bemorning vazni 200 kg
Og'irligi 40 kg